



Solventum MedTech OEM

Product clinical data summary

Product Name: Solventum™ Medical Tape 9841

Effective: June 1997

Medical Tape 9841 has been subjected to the following safety evaluations:

In Vitro Cytotoxicity (Agar Overlay)

Protocol reference: Guess, W. L. et al; "Agar Diffusion Method for Toxicity Screening of Plastics on Cultured Cell Monolayers" J. Pharm. Sci. 54:1545-1547 (1965).

Results: 0.0/0.0

Acute Primary Skin Irritation on Albino Rabbits

Protocol reference: Draize: Appraisal of the Safety of Chemicals in Food, Drugs and Cosmetics (1965). Published by the Editorial Committee of the Association of Food and Drug Officials of the United States.

Results: 1.2/8.0.

Acute Intracutaneous Irritation in Albino Rabbits

Protocol reference: Draize: Appraisal of the Safety of Chemicals in Food, Drugs and Cosmetics (1965). Published by the Editorial Committee of the Association of Food and Drug Officials of the United States.

Results: No leachable toxic substances.

Acute Systemic Toxicity in Albino Mice

Protocol reference: U.S. Pharmacopoeia XXII, 1990, pg. 1499.

Results: No toxic leachables.

Repeated Insult Patch Test (Draize) in Humans

Protocol reference: Draize: Appraisal of the Safety of Chemicals in Food, Drugs and Cosmetics (1965). Published by the Editorial Committee of the Association of Food and Drug Officials of the United States.

Results: No subjects showed evidence of sensitization. Low potential for sensitization as demonstrated in this evaluation.

21 Day Cumulative Irritation Test in Humans

Protocol reference: Draize: Appraisal of the Safety of Chemicals in Food, Drugs and Cosmetics (1965). Published by the Editorial Committee of the Association of Food and Drug Officials of the United States.

Results: No untoward effects were observed with any of the subjects entered in this study.

These tests are in accordance with the ISO 10933 Part-1 "Biological Evaluation of Medical Devices", as put forth by the FDA. Product Number 9841 has satisfied the requirements for devices in contact with intact skin for short term application (up to 29 days).

It is the responsibility of our customers to determine final suitability of our products for their application.

Solventum MedTech OEM

Phone 1-800-742-1994

Web go.Solventum.com/MedTechOEM

© Solventum 2025. Solventum, the S logo and other trademarks are trademarks of Solventum or its affiliates. Other trademarks are the property of their respective owners.